

Manufacturer Disclosure Statement for Medical Device Security – MDS²

SECTION 1

Device Category	Manufacturer	Document ID	Document Release Date
Class II	Vital Images, Inc.	VLC-09115	2-Jun-16
Device Model	Software Revision	Software Release Date	
Vitrea	6.8	8-Jun-16	
Manufacturer or Representative Contact Information:	Company Name	Manufacturer Contact Information	
	Representative Name/Position	5850 Opus Parkway Suite #300, Minnetonka MN 55343	
	Kelly Dupasquier Product Manager	952-487-9500	

MANAGEMENT OF ELECTRONIC PROTECTED HEALTH INFORMATION (ePHI)	Yes	No	N/A	Note #
1. Can this device transmit or maintain electronic Protected Health Information (ePHI)?	Yes			2
2. Types of ePHI data elements that can be maintained by the device:				
a. Demographic (e.g., name, address, location, unique identification number)?	Yes			_____
b. Medical record (e.g., medical record #, account #, test or treatment date, device identification number)?	Yes			_____
c. Diagnostic/therapeutic (e.g., photo/radiograph, test results, or physiologic data with identifying characteristics)?	Yes			_____
d. Open, unstructured text entered by device user/operator?	Yes			_____
3. Maintaining ePHI - Can the device				
a. Maintain ePHI temporarily in volatile memory (i.e., until cleared on by power-off or reset)?	Yes			_____
b. Store ePHI persistently on local media?	Yes			_____
c. Import/export ePHI with other systems?	Yes			_____
4. Mechanisms used for the transmitting, importing/exporting of ePHI – Can the device				
a. Display ePHI (e.g., video display)?	Yes			_____
b. Generate hardcopy reports or images containing ePHI?	Yes			_____
c. Retrieve ePHI from or record ePHI to removable media (e.g., disk, DVD, CD-ROM, tape, CF/SD card, memory stick)?	Yes			_____
d. Transmit/receive or import/export ePHI via dedicated cable connection (e.g., IEEE 1073, serial port, USB, FireWire)?	Yes			_____
e. Transmit/receive ePHI via a network connection (e.g., LAN, WAN, VPN, intranet, Internet)?	Yes			_____
f. Transmit/receive ePHI via an integrated wireless connection (e.g. WiFi, Bluetooth, infrared)?	Yes			_____
g. Other?			N/A	_____

ADMINISTRATIVE SAFEGUARDS	Yes	No	N/A	Note #
5. Does manufacturer offer operator and technical support training or documentation on device security features?	Yes			_____
6. What underlying operating system(s) (including version number) are used by the device?				_____
Microsoft Windows Server 2008 R2 SP1 Standard x64 Edition (servers only)				
Microsoft Windows Server 2012 R2 Standard x64 Edition (servers only)				
Microsoft Windows 7 Professional 32-bit Edition with SP1 (clients only)				
Microsoft Windows 7 Professional 64-bit Edition with SP1 (clients and workstations)				
Microsoft Windows 8.1 Professional 64-bit Edition (clients and workstations)				

PHYSICAL SAFEGUARDS	Yes	No	N/A	Note #
7. Are all device components maintaining ePHI (other than removable media) physically secure (i.e. cannot remove without tools)?			N/A	3
8. Does the device have an integral data backup capability (i.e., backup onto removable media like tape, disk)?			N/A	3
9. Can the device boot from uncontrolled or removable media (i.e., a source other than an internal drive or memory component)?			N/A	3

TECHNICAL SAFEGUARDS	Yes	No	N/A	Note #
10. Can software or hardware not authorized by the device manufacturer be installed on the device without the use of tools?			N/A	3
11. Can the device be serviced remotely (i.e., maintenance activities performed by service person via network or remote connection)?	Yes			4
a. Can the device restrict remote access to specific devices or network locations (e.g., specific IP addresses)?			N/A	3
b. Can the device provide an audit trail of remote-service activity?			N/A	3
c. Can security patches or other software be installed remotely?			N/A	3
12. Level of owner/operator service access to device operating system: Can the device owner/operator				
a. Apply device manufacturer-validated security patches?			N/A	5
b. Install or update antivirus software?	Yes			3
c. Update virus definitions on manufacturer-installed antivirus software?	Yes			5
d. Obtain administrative privileges (e.g. access operating system or application via local root or admin account)?	Yes			3
13. Does the device support user/operator specific username and password?	Yes			_____
14. Does the system force reauthorization after a predetermined length of inactivity (e.g., auto login, session lock)?			N/A	3

Manufacturer Disclosure Statement for Medical Device Security – MDS²

SECTION 1

Device Category	Manufacturer	Document ID	Document Release Date
Class II	Vital Images, Inc.	VLC-09115	2-Jun-16
Device Model	Software Revision	Software Release Date	
Vitrea	6.8	8-Jun-16	
Manufacturer or Representative Contact Information:	Company Name	Manufacturer Contact Information	
	Vital Images, Inc.	5850 Opus Parkway Suite #300, Minnetonka MN 55343	
	Representative Name/Position	952-487-9500	
	Kelly Dupasquier Product Manager		

15. Events recorded in device audit trail (e.g., user, date/time, action taken): Can the audit trail record.....		
a. Login and logout by users/operators?.....	Yes	2
b. Viewing of ePHI?.....	Yes	2
c. Creation, modification or deletion of ePHI?.....	Yes	2
d. Import/export or transmittal/receipt of ePHI?.....	Yes	2
16. Does the device incorporate an emergency access ("break-glass") feature that is logged?.....	No	3
17. Can the device maintain ePHI during power service interruptions?.....	Yes	3
18. Controls when exchanging ePHI with other devices:.....		
a. Transmitted only via a point-to-point dedicated cable?.....	N/A	3
b. Encrypted prior to transmission via a network or removable media?.....	Yes	3
c. Restricted to a fixed list of network destinations.....	Yes	3
19. Does the device ensure the integrity of the ePHI data with implicit or explicit error detection/correction technology?.....	No	3

Other Security Considerations

No additional security concerns identified.

Manufacturer Disclosure Statement for Medical Device Security – MDS²

SECTION 1

Device Category	Manufacturer	Document ID	Document Release Date
Class II	Vital Images, Inc.	VLC-09115	2-Jun-16
Device Model	Software Revision	Software Release Date	
Vitrea	6.8	8-Jun-16	
Manufacturer or Representative Contact Information:	Company Name	Manufacturer Contact Information	
	Representative Name/Position	5850 Opus Parkway Suite #300, Minnetonka MN 55343	
	Kelly Dupasquier Product Manager	952-487-9500	

SECTION 2

EXPLANATORY NOTES (from questions 1 - 19)

IMPORTANT: Refer to Section 2.2.2 of this standard for the proper interpretation of information requested in this form

1. The term "device" herein refers to the VitreaAdvanced[®] software and /or license. Vital Images does not manufacture hardware or operating system software and /or components.
2. The VitreaAdvanced[®] software can create audit logs that list user actions to ePHI data. ePHI information available to users is in the form of DICOM data and Reports. Any time a user loads a dataset, creates snapshots or batches, exports, reports on, prints, burns to media or deletes it from VitreaAdvanced, action is logged in an audit log file
3. Vital Images, Inc. does not manufacture hardware or operating system software and /or components which may extend these functions.
4. Vital Images, Inc. Technical Support can use a remote access tool for remote application (Vitrea[®]) support
5. Vital Images, Inc. does not provide customers with Anti-Virus Definition Files, Operating System Security or Service Packs. Customer is responsible for managing hardware security and software update policies.